



test parameters	column: CHIROBIOTIC V, 25 cm x 4.6 mm I.D., 5 µm particles (11024AST)
	mobile phase: (A) methanol; (B) 1% triethylamine, pH 4/1; (95:5, A:B)
	flow rate: 1 mL/min
	column temp.: 25 °C
	detector: UV, 220 nm
	injection: 2 µL
	sample: 1 mg/mL in methanol