



test parameters	column: CHIROBIOTIC T, 25 cm x 4.6 mm I.D., 5 µm particles (12024AST)
	mobile phase: (A) 0.1% triethylamine acetate, pH 3.8; (B) ethanol; (20:80, A:B)
	flow rate: 1 mL/min
	column temp.: 23 °C
	detector: UV, 225 nm
	injection: 5 µL
	sample: 5 mg/mL in methanol