



Figure used with permission of Mitsubishi Chemical Co.

test parameters	<p>column: Discovery BIO PolyMA-SCX, 5 cm × 4.6 mm I.D., 5 μm particles (56901-U)</p> <p>mobile phase: (A) 20 mM Bis-Tris HCl, pH 7.0; (B) A + 0.5 M NaCl 24-69% B over 20 min</p>
	flow rate: 0.5 mL/min
	detector: UV, 280 nm
	injection: 10 μg each variant