



test parameters	column: SUPELCOSIL LC-DABS, 15 cm × 4.6 mm I.D., 3 µm particles (59137)
	mobile phase: (A) 0.025 M KH ₂ PO ₄ , pH 7.0 with potassium hydroxide; (B) acetonitrile:methanol; (70:30, A:B)
	flow rate: 1.5 mL/min
	detector: VIS, 436 nm
	injection: 5 µL; 50pmole each amino acid on-column