

# Verzele test

Column: 50 x 2.1 mm ID

Mobile phase: MeOH:0.5% NaAc (pH 7.8) = 60:40

Flow rate: 0.3 ml/min

Detection: UV 254 nm

Sample: 1. Acetylaceton  
2. 1-Nitronaphthalene  
3. Naphthalene

