



Column	TCI Dual ODS-CX15 4.6 mmI.D. X 150mm
Mobile Phase	A) CH ₃ CN / 50mmol/L HCOONH ₄ (pH3.5,HCOOH) = 10/90B) CH ₃ CN / 50mmol/L HCOONH ₄ (pH3.5,HCOOH) = 70/30A to 50% B in 30min
Temperature	25 °C
Detection	UV 254 nm
Flow Rate	1.0 mL/min
Sample	1.5-Hydroxy-L-tryptophan 2.5-Hydroxyindole-3-acetic Acid(5-HIAA) 3.Acetylserotonin 4.L-Tryptophan 5.Serotonin(5HT) 6.Melatonin