Gliotoxin from *Aspergillus Fumigatus* Liquid Culture

ACE®
Ultra-inert
UHPLC & HPLC Columns


Application #AN3780


Conditions

Column:	ACE 5 C18
Dimensions:	250 x 4.6 mm
Part Number:	ACE-121-2546
Mobile Phase:	A: 0.05% TFA in H_2O/MeCN (90:10 v/v)
	B: 0.05% TFA in H_2O/MeCN (40:60 v/v)

Time (mins)	%B
0	10
21	100

Flow Rate:	1 mL/min
Detection:	UV, 254 nm





1. Gliotoxin

Svahn KS, Goransson U, Chryssanthou E, Olsen B, Sjolin J, Stromstedt A. Induction of Gliotoxin Secretion in Aspergillus fumigatus by Bacteria-Associated Molecules. PloS ONE 9(4): e93685. doi:10.1371/journal.pone.0093685

© 2015 ACT Ltd. All rights reserved. All products are available worldwide.