



test parameters	column: CHIROBIOTIC R, 15 cm x 2.1 mm, 5 μm particles (13019AST)
	mobile phase: (A) 20 mM ammonium acetate, pH 5.6 (B) methanol; (70:30, A:B)
	flow rate: 0.2 mL/min
	column temp.: 35 °C
	detector: ESI(-)