



test parameters	<b>column:</b> CHIROBIOTIC R, 15 cm x 4.6 mm I.D., 5 $\mu\text{m}$ particles (13023AST)
	<b>mobile phase:</b> (A) 20 mM ammonium acetate, pH 5.6; (B) methanol; (50:50, A:B)
	<b>flow rate:</b> 0.6 mL/min
	<b>column temp.:</b> 25 °C
	<b>detector:</b> UV, 254 nm
	<b>injection:</b> 5 $\mu\text{L}$
	<b>sample:</b> 1 mg/mL in methanol