



test parameters	column: CYCLOBOND I 2000 HP-RSP, 25 cm × 4.6 mm I.D., 5 μm particles (24024AST)
	mobile phase: (A) 10 mM ammonium acetate, pH 4.0; (B) acetonitrile; (80:20, A:B)
	flow rate: 1 mL/min
	column temp.: 23 °C
	detector: UV, 235 nm
	injection: 5 μL