



test parameters	column: Astec CLC-D, 15 cm x 4.6 mm I.D., 5 µm particles (53023AST)
	mobile phase: 3 mM copper sulfate, pH 3.2
	flow rate: 1 mL/min
	column temp.: 25 °C
	detector: UV, 254 nm
	injection: 10 µL
	sample: 5 mg/mL in water