



test parameters	<p>column: Astec CLC-L, 15 cm x 4.6 mm I.D., 5 μm particles (53123AST)</p> <p>mobile phase: 5 mM copper sulfate</p>
	flow rate: 1 mL/min
	column temp.: 25 °C
	detector: UV, 254 nm
	injection: 10 μL
	sample: 5 mg/mL in water