



test parameters	column: Ascentis Express RP-Amide, 10 cm x 3 mm I.D., 2.7 $\mu$ m particles (53918-U)
	mobile phase: (A) water; (B) methanol; (60:40, A:B)
	flow rate: 0.6 mL/min
	pressure: 200 bar
	column temp.: 50 °C
	detector: UV, 250 nm
	injection: 1 $\mu$ L