



test parameters	column: Ascentis Express RP-Amide, 10 cm x 4.6 mm I.D., 2.7 µm particles (53929-U)
	mobile phase: (A) 20 mM phosphoric acid, pH 2; (B) acetonitrile; (75:25, A:B)
	flow rate: 1.5 mL/min
	column temp.: 35 °C
	detector: UV, 270 nm
	injection: 25 µL
	sample: 50 mg/L in mobile phase