



test parameters	column: Ascentis Express RP-Amide, 15 cm x 4.6 mm I.D., 2.7 $\mu$ m particles (53931-U)
	mobile phase: methanol
	flow rate: 1 mL/min
	pressure: 114 bar
	column temp.: 30 $^{\circ}$ C
	detector: UV-VIS, 454 nm
	injection: 5 $\mu$ L
	sample: V8 Splash <sup>®</sup> fruit juice was extracted with hexane, the extract dissolved 1:1 into ethanol for HPLC injection