



test parameters	column: Discovery HS F5, 15 cm × 4.6 mm A.D., 5 µm particles (567516-U)
	mobile phase: (A) water (10 mM ammonium sulfate, pH 3.0 with formic acid): (B)acetonitrile; (80:20, A:B)
	flow rate: 1.0 mL/min
	column temp.: 35 °C
	detector: UV, 254 nm
	injection: 10 µL
	sample: 50 µg/mL each