



test parameters	column: Ascentis Express C8, 5 cm × 4.6 mm I.D., 2.7 µm particles (58336-U)
	mobile phase: (A) water; (B) acetonitrile; (45:55, A:B)
	flow rate: 1.76 mL/min
	column temp.: 30 °C
	detector: UV, 254 nm
	injection: 5 µL